



For Laboratory Instruments

# FILIPS 400

Semi-Auto Biochemistry Analyzer



Qr



## Technical Parameters

Light Source : Halogen Lamp 6 V, 10 W

Screen Touch , 7inch , Keyboard , Mouse ,USB

Suction System & Cuvette System. Injection: 0~6000ul

Interface : Standard Rs-232 interface , 4 USB interface,

## Exclusive distributor



Add : Zohour District, Zagazig 14512 , Egypt .

Tel efax : 0552329695

Mob : 01220108221 - 01024025659 - 01557468221

E-mail : support@guda-tek.com

Website : [www.guda-tek.com](http://www.guda-tek.com)



# FILIPS 400

For Laboratory Instruments Semi-Auto Biochemistry Analyzer



## Introduction

Multiple measurement and multiple calculation methods

Store 500 test items and 10000 test results

Each item can be set for two quality-control

Self-check function for alarming to light route etc

With standard interface for communication

convenient for data transmitting and processing

The Semi-auto Biochemistry Analyzer Measure

Biochemical Indes by analyzing

body fluid , then combines With other clinical

information,to help diagnose disease,evaluate,

organs function,identify disease gene

determine the norm for future therapy .

